

LMS  
mmB

## INTOX EC/IR II Quality Assurance Worksheet

Instrument Serial Number 010448 Worksheet Start Date 11/13/2014  
Location Virginia Beach Jail  
Address 2501 James Madison Blvd., Virginia Beach, VA 23456  
DFS Technician Heather Stanton License No. 18910

☐ Laboratory

☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) 761 Reference Barometer (mm HG) 762  
Reference Barometer(RB)Serial # 009113 RB Calibration Due 4/28/2015

### Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.300		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.100		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Dry gas standard Lot No. (with tank no.)

AG419702-33

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☐ Removed to DFS-Central

Supplies  
Mouthpieces  
Certificates of Analysis  
Operator Worksheet  
Other:



Notes:

Instrument Serial Number

010448

Certification Date

☐ Calibrated☐ Certified☐ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section

DFS Technician

Heath M. Stash

Date

11/14/14

Issuing Analyst

Michelle Bullitt

Date

11/14/141PD




COMMONWEALTH OF VIRGINIA  
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS  
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

lines  
mmb

NAME OF ACCUSED <b>INSTRUMENT, TEST,</b>	NAME OF COURT <b>DFS</b>
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY <b>STANTON, HEATHER, M</b>		AGENCY <b>DFS Central Lab</b>
DFS LICENSE NUMBER <b>18910</b>	LICENSE EXPIRES <b>10/01/2016</b>	DATE TEST CONDUCTED <b>11/13/2014</b>
TEST EQUIPMENT NUMBER <b>010448</b>		

RESULTS: TIME SAMPLE TAKEN 13:28 EST

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS \_\_\_\_\_ DAY OF \_\_\_\_\_, 20 \_\_\_\_\_.

BREATH TEST OPERATOR
----------------------

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS \_\_\_\_\_  
SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS \_\_\_\_\_  
OPERATOR'S SIGNATURE



# IntoxNet MIS Report

Report Generated 14 Nov 2014 at 09:41

LMS  
mmb

## Test Results

Instrument Serial Number 010448

Test # 001408 Subject Test

Test Location 1 Department of  
Test Date 13 Nov 2014

Test Location 2 Forensic Science  
Test Time 13:22  
Remote/Local Local

Test Location 3  
System Check Passed

Operator's Last Name STANTON  
Agency DFS Central Lab

Operator's First Name HEATHER

Operator's Middle Initial M

Card Serial Number 118910

License Number 18910

Effective Date 10/01/2014

Expiration Date 10/01/2016

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 13 Nov 2014 End Time 13:29

Result Time 13:28

Result Date 13 Nov 2014

Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 13:22

Data Type BLK

Sample Value 0.000

Sample Time 13:23

Data Type CHK

Sample Value 0.101

Sample Time 13:23

Data Type BLK

Sample Value 0.000

Sample Time 13:25

Data Type SUBJ

Sample Value 0.000

Sample Time 13:25

Data Type BLK

Sample Value 0.000

Sample Time 13:26

Data Type SUBJ

Sample Value 0.000

Sample Time 13:28

Data Type BLK

Sample Value 0.000

Sample Time 13:29

Standard Type Dry Gas Std

Standard Value 0.100

Standard Lot Number AG419702-33

Standard Expiration Date 07/16/2016

Tank Pressure 850

Barometric Pressure 761 mmHg

Blow Sample Number 1 Blow Duration 3.56 sec

Blow Volume 1821 cc

End-of-Blow Time 13:25

Blow Sample Number 2 Blow Duration 3.50 sec

Blow Volume 1729 cc

End-of-Blow Time 13:28

Tamper Evident Stamp ffa5afe6

Test Status Code 0

Test Status Success